

Test Report Efficiency of S.C.A.T. SafetyCaps



HPLC Safety Set

-

Problem

What happens if an HPLC user just guides the mobile phase through capillaries in "open caps", instead of using S.C.A.T. SafetyCaps?

Test chromatograms of 3 PAHs (polycyclic aromatic hydocarbons) were used for a 31-day comparison.

Procedure

All 4 bottles were filled at the beginning with the identical mixture of water + methanol = 20 + 80 (percentage by weight). Using bottle B as a reference, a chromatogram of a mixture of three PAHs (polycyclic aromatic hydrocarbons) - naphthalene, pyrene and chrysene - was acquired for comparison. After the reference chromatogram was recorded, all bottles were stored at room temperature in a fume hood, which guaranteed a gentle air flow over the top of the bottles, for 31 days.





Test Conditions

Bottle A: This bottle was closed using a S.C.A.T. SafetyCap with a GL-45 thread. (Part No. 107 019)









Bottle C: This bottle was closed using a cap with a 10-mm hole in the plastic material, yielding an open exit area of approximately 0.785 cm².



Bottle D: This bottle was closed using a cap with three 3-mm holes in the plastic material, yielding an open exit area of approximately 0.212 cm².



HPLC-System: VWR HITACHI LaChrom Elite[®] system with Diode Array Detector, EZChrom Elite[™] Software, Isocratic pump conditions and premixed mobile phase.

HPLC-Column: Purospher[®] RP-18e (5 μ), 125 x 4 mm.



Test Results

The test clearly revealed that unless closed solvent delivery systems - like the ones guaranteed by using the S.C.A.T. SafetyCaps - are used, unreliable retention times may be encountered even after a relatively short time.

As expected, **bottles A and B** did not show any significant changes in their weights, so no solvent vapors escaped from those two bottles. In contrast, bottles C and D showed significant and uncontrolled loss of liquid through evaporation (see chart below). After 31 days, the separation of the three PAHs was repeated under identical HPLC conditions (same HPLC system, same column, etc.) using the liquids from bottles C and D. The result was a significant prolongation of all retention times of the test compounds, which would make compound identification based on retention times impossible. Assuming a more or less linear relationship between evaporation of the mobile phase and time, it becomes clear that even after one day of using mobile phase bottles that are not tightly closed, the user can expect changes in retention times.

Weight loss of bottles over a period of 31 days:



	L			
	Bottle A	Bottle B	Bottle C	Bottle D
Day 1	457.45	539.26	724.14	715.08
Day 31	457.43	539.26	672.45	687.36
Loss (g)	0.02	0.00	51.69	27.72
Loss (%)	0.004	0.000	7.138	3.876





Conclusion

If tightly sealed reservoir bottles for mobile phases are not used, there is a clear risk of uncontrolled vaporization, possibly resulting in wrong retention time assignment for eluting compounds. The use of S.C.A.T. SafetyCaps definitely prevents such uncontrolled evaporation, and in addition protects your laboratory air from constant contamination with toxic compounds!





S.C.A.T. Europe GmbH Opelstraße 3 64546 Mörfelden Germany

 Phone:
 +49 - (0) 6105 - 30 55 86 - 0

 Fax:
 +49 - (0) 6105 - 30 55 86 - 99

 e-Mail:
 info@scat-europe.com

 Web:
 www.scat-europe.com

